







## **CERTIFICATE**

## The ATP-Test Station Of the TÜV SÜD Industrie Service GmbH

confirms to the company

Comet System, s.r.o 1.maje 1220 CZE-75661 Roznov pod Radhostem

because of the positive results of the tested temperature recorder,

G0241

according to the test-report No. 1086 TS from the 14th of December 2011, that the temperature recorder fulfils the requirements of the standard DIN EN 12830 for the determination:

Accuracy class: 1

environmental conditions A

application range: transportation

The manufacturer is allowed to mark the temperature recorder with the following test-report number:

ATP-MUC 1086 TS

This certificate

is valid from

14.12.2011

This certificate

is valid to:

14.12.2016

ATP-Prüfstelle

Rudolf Glück

Munich, the 14.12.2011



ATP

Technical expert Thorsten Bennert

TÜV®