



Calibration Certificate

No. 11962032/001

Model T7411
Serial Number 11962032
Number of pages 1

Process of Calibration Direct comparison

Conditions of Measurement

| | | |
|-------------------|-------|--------|
| Temperature | 23 °C | ± 3 °C |
| Relative Humidity | 40 % | ± 20 % |

Calibration Standards (valid to)

Thermometer F200 (PE136), 008408/01+J0295A-1-4, (20.5.2012)

Pressure meter DPI142, PE139, 2421692, (30.6.2011)

Humidity meter Testo 645, PE140, 20145248/801, (21.7.2011)

All standards are traceable to ČMI (Czech Metrology Institute, a signatory to the arrangement CIPM MRA, see www.bipm.org) or in terms of relative humidity to DKD accredited laboratory Testo Germany.

Measuring Results

Channel 1-Relative Humidity

| Standard | Device | Uncertainty | Used Standard | Note |
|----------|--------|-------------|---------------|------|
| 49.8 % | 49.8 % | 1.3 % | PE140 | |

Channel 2-Pressure

| Standard | Device | Uncertainty | Used Standard | Note |
|------------|-----------|-------------|---------------|------|
| 756.34 hPa | 756.3 hPa | 0.28 hPa | PE139 | |
| 972.83 hPa | 972.7 hPa | 0.28 hPa | PE139 | |

Channel 3-Temperature

| Standard | Device | Uncertainty | Used Standard | Note |
|----------|---------|-------------|---------------|------|
| 22.28 °C | 22.3 °C | 0.14 °C | PE136 | |

Measuring Uncertainty

The expanded uncertainty of measurement corresponding to the measurement results is started as the standard uncertainty of measurement multiplied by the coverage factor $k=2$. Usually the values is located in the corresponding interval with probability of approximately 95%. This was determined in accordance with EA4/02.

Date of Calibration

21.6.2011

Calibrated By

.....
Jaroslava Kubáňová

Has Approved

.....
Bc. Petr Krčmář